



Welch Allyn®
WAM™/UTK™

Wireless Acquisition Module User Manual



Manufactured by Welch Allyn, Inc., Skaneateles Falls, NY U.S.A



CAUTION: Federal law restricts this device to sale by or on the order of a physician

© 2020 This document contains confidential information that belongs to Welch Allyn, Inc. No part of this document may be transmitted, reproduced, used, or disclosed outside of the receiving organization without the express written consent of Welch Allyn, Inc. Welch Allyn is a registered trademark of Welch Allyn, Inc. ELI and WAM are trademarks of Welch Allyn, Inc.

Software V2.1X

The information in this document is subject to change without notice.

PATENT/PATENTS

hillrom.com/patents

May be covered by one or more patents. See above Internet address. The Hill-Rom companies are the proprietors of European, US, and other patents and pending patent applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at 1.888.667.8272, mor_tech.support@hillrom.com.

REF 9515-174-50-ENG Rev M
Revision Date: 2020-06

901095 ECG ACQUISITION MODULE



Welch Allyn, Inc.
4341 State Street Road
Skaneateles Falls, NY 13153 USA

EC REP and EU IMPORTER

Welch Allyn Limited
Navan Business Park, Dublin Road,
Navan, Co. Meath C15 AW22
Ireland

Authorized Australian Sponsor

Welch Allyn Australia Pty. Ltd.
Unit 4.01, 2-4 Lyonpark Road
Macquarie Park NSW 2113
Phone 800 650 083

hillrom.com

Welch Allyn, Inc. is a subsidiary of Hill-Rom Holdings, Inc.



TABLE OF CONTENTS

NOTICES	3
MANUFACTURER’S RESPONSIBILITY	3
RESPONSIBILITY OF THE CUSTOMER	3
EQUIPMENT IDENTIFICATION.....	3
COPYRIGHT AND TRADEMARK NOTICES.....	3
OTHER IMPORTANT INFORMATION.....	3
NOTICE TO EU USERS AND/OR PATIENTS	3
WARRANTY INFORMATION	4
YOUR WELCH ALLYN WARRANTY.....	4
USER SAFETY INFORMATION	6
WARNING	6
CAUTION:	8
NOTES:.....	9
EQUIPMENT SYMBOLS AND MARKINGS	10
SYMBOL DELINEATION	10
PACKAGE SYMBOL DELINEATION.....	12
GENERAL CARE	13
PRECAUTIONS	13
INSPECTION	13
CLEANING AND DISINFECTING	13
DISPOSAL	14
INTRODUCTION	15
MANUAL PURPOSE	15
AUDIENCE	15
SYSTEM DESCRIPTION.....	15
INDICATIONS FOR USE	15
WAM™ WIRELESS ACQUISITION MODULE AND UTK RECEIVER	15
WAM WITH LEAD WIRES	16
USING THE BUTTONS.....	16
APPROVED BATTERY MODELS	16
PART NUMBERS	16
WAM SPECIFICATIONS.....	17
UTK SPECIFICATIONS	19
EQUIPMENT PREPARATION	20
BATTERY INSTALLATION	20
APPLYING POWER.....	20
ATTACHING THE LEAD WIRE CONNECTOR BLOCK	20
LABELING THE WAM AND ELECTROCARDIOGRAPH.....	21
PAIRING WITH AN ELI 150c/ELI 250c.....	21
PAIRING WITH AN ELI 230	21
PAIRING WITH AN ELI 280	21
PAIRING WITH AN ELI 350	22

PAIRING WITH AN ELI 38022

PAIRING THE WAM WITH Q-STRESS23

PAIRING THE WAM WITH XSCRIBE23

WAM UTK COMPATIBILITY24

LEAD FAIL24

LED INDICATORS24

ACQUIRING AN ECG25

MAINTENANCE26

 PERIODIC MAINTENANCE26

MESSAGES AND INFORMATION27

 SYSTEM INFORMATION LOG28

 SERIAL AND PART NUMBER LOCATION28

ELECTROMAGNETIC COMPATIBILITY (EMC)29

 GUIDANCE AND MANUFACTURER’S DECLARATION: ELECTROMAGNETIC EMISSIONS30

 GUIDANCE AND MANUFACTURER’S DECLARATION: ELECTROMAGNETIC IMMUNITY.....30

 GUIDANCE AND MANUFACTURER’S DECLARATION: ELECTROMAGNETIC IMMUNITY.....31

 REGULATORY RADIO COMPLIANCE33

 INDUSTRY CANADA (IC) EMISSIONS.....34

 EUROPEAN UNION35

NOTICES

Manufacturer's Responsibility

Welch Allyn, Inc. is responsible for the effects on safety and performance only if:

- Assembly operations, extensions, readjustments, modifications, or repairs are carried out only by persons authorized by Welch Allyn, Inc.
- The device is used in accordance with the instructions for use.

Responsibility of the Customer

The user of this device is responsible for ensuring the implementation of a satisfactory maintenance schedule. Failure to do so may cause undue failure and possible health hazards.

Equipment Identification

Welch Allyn, Inc. equipment is identified by a serial and reference number on the back of the device. Care should be taken so that these numbers are not defaced.

The product label is applied showing unique identification numbers along with other important information printed on the label.

The serial number format is as follows:

YYYWWSSSSSS

YYY = First Y is always 1 followed by two-digit Year of manufacture

WW = Week of manufacture

SSSSSS = Sequence number of manufacture

When present, the external UTK label will include a reference number (REF) and a lot number. The UDI label (when applicable) is placed below the product label. This label is placed to the right of the product label.

Copyright and Trademark Notices

This document contains information that is protected by copyright. All rights are reserved. No part of this document may be photocopied, reproduced, or translated to another language without prior written consent of Welch Allyn, Inc.

Other Important Information

The information in this document is subject to change without notice.

Welch Allyn, Inc. makes no warranty of any kind with regard to this material including, but not limited to, implied warranties of merchantability and fitness for a particular purpose. Welch Allyn, Inc. assumes no responsibility for any errors or omissions that may appear in this document. Welch Allyn, Inc. makes no commitment to update or to keep current the information contained in this document.

Notice to EU Users and/or Patients

Any serious incident that has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

WARRANTY INFORMATION

Your Welch Allyn Warranty

WELCH ALLYN, INC. (hereafter referred to as “Welch Allyn”) warrants that components within Welch Allyn products (hereafter referred to as “Product/s”) will be free from defects in workmanship and materials for the number of years specified on documentation accompanying the product, or previously agreed to by the purchaser and Welch Allyn, or if not otherwise noted, for a period of twelve (12) months from the date of shipment.

Consumable, disposable or single use products such as, but not limited to, PAPER or ELECTRODES are warranted to be free from defects in workmanship and materials for a period of 90 days from the date of shipment or the date of first use, whichever is sooner.

Reusable product such as, but not limited to, BATTERIES, BLOOD PRESSURE CUFFS, BLOOD PRESSURE HOSES, TRANSDUCER CABLES, Y-CABLES, PATIENT CABLES, LEAD WIRES, MAGNETIC STORAGE MEDIUMS, CARRY CASES or MOUNTS, are warranted to be free from defects in workmanship and materials for a period of 90 days. This warranty does not apply to damage to the Product/s caused by any or all of the following circumstances or conditions:

- a) Freight damage;
- b) Parts and/or accessories of the Product/s not obtained from or approved by Welch Allyn;
- c) Misapplication, misuse, abuse, and/or failure to follow the Product/s instruction sheets and/or information guides;
- d) Accident; a disaster affecting the Product/s;
- e) Alterations and/or modifications to the Product/s not authorized by Welch Allyn;
- f) Other events outside of Welch Allyn’s reasonable control or not arising under normal operating conditions.

THE REMEDY UNDER THIS WARRANTY IS LIMITED TO THE REPAIR OR REPLACEMENT WITHOUT CHARGE FOR LABOR OR MATERIALS, OR ANY PRODUCT/S FOUND UPON EXAMINATION BY WELCH ALLYN TO HAVE BEEN DEFECTIVE. This remedy shall be conditioned upon receipt of notice by Welch Allyn of any alleged defects promptly after discovery thereof within the warranty period. Welch Allyn’s obligations under the foregoing warranty will further be conditioned upon the assumption by the purchaser of the Product/s (i) of all carrier charges with respect to any Product/s returned to Welch Allyn’s principal place or any other place as specifically designated by Welch Allyn or an authorized distributor or representative of Welch Allyn, and (ii) all risk of loss in transit. It is expressly agreed that the liability of Welch Allyn is limited and that Welch Allyn does not function as an insurer. A purchaser of a Product/s, by its acceptance and purchase thereof, acknowledges and agrees that Welch Allyn is not liable for loss, harm, or damage due directly or indirectly to an occurrence or consequence therefrom relating to the Product/s. If Welch Allyn should be found liable to anyone under any theory (except the expressed warranty set forth herein) for loss, harm, or damage, the liability of Welch Allyn shall be limited to the lesser of the actual loss, harm, or damage, or the original purchase price of the Product/s when sold.

EXCEPT AS SET FORTH HEREIN WITH RESPECT TO REIMBURSEMENT OF LABOR CHARGES, A PURCHASER'S SOLE EXCLUSIVE REMEDY AGAINST WELCH ALLYN FOR CLAIMS RELATING TO THE PRODUCT/S FOR ANY AND ALL LOSSES AND DAMAGES RESULTING FROM ANY CAUSE SHALL BE THE REPAIR OR REPLACEMENT OF DEFECTIVE PRODUCT/S TO THE EXTENT THAT THE DEFECT IS NOTICED AND WELCH ALLYN IS NOTIFIED WITHIN THE WARRANTY PERIOD. IN NO EVENT, INCLUDING THE CLAIM FOR NEGLIGENCE, SHALL WELCH ALLYN BE LIABLE FOR INCIDENTAL, SPECIAL, OR CONSEQUENTIAL DAMAGES, OR FOR ANY OTHER LOSS, DAMAGE, OR EXPENSE OF ANY KIND, INCLUDING LOSS OF PROFITS, WHETHER UNDER TORT, NEGLIGENCE OR STRICT LIABILITY THEORIES OF LAW, OR OTHERWISE. THIS WARRANTY IS EXPRESSLY IN LIEU OF ANY OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO THE IMPLIED WARRANTY OF MERCHANTABILITY AND THE WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE.

USER SAFETY INFORMATION



WARNING: Means there is the possibility of personal injury to you or others.



Caution: Means there is the possibility of damage to the device.

Note: Provides information to further assist in the use of the device.



WARNING

- This manual gives important information about the use and safety of this device. Deviating from operating procedures, misuse or misapplication of the device, or ignoring specifications and recommendations could result in increased risk of harm to users, patients and bystanders, or damage to the device.
- Device transmits data reflecting a patient's physiological condition to a properly equipped receiving device that when reviewed by a trained physician or clinician can be useful in determining a diagnosis; however, the data should not be used as a sole means for determining a patient's diagnosis.
- Users are expected to be licensed clinical professionals knowledgeable about medical procedures and patient care, and adequately trained in the use of this device. Before attempting to use this device for clinical applications, the operator must read and understand the contents of the user manual and other accompanying documents. Inadequate knowledge or training could result in increased risk of harm to users, patients and bystanders, or damage to the device. Contact Welch Allyn service for additional training options.
- To maintain designed operator and patient safety, peripheral equipment and accessories used that can come in direct patient contact must be in compliance with UL 60601-1, IEC 60601-1, and IEC 60601-2-25. Only use parts and accessories supplied with the device and available through Welch Allyn, Inc.
- Patient cables intended for use with the device include series resistance (9 Kohm minimum) in each lead for defibrillation protection. Patient cables should be checked for cracks or breakage prior to use.
- Conductive parts of the patient cable, electrodes, and associated connections of type CF applied parts, including the neutral conductor of the patient cable and electrodes, should not come into contact with other conductive parts including earth ground.
- ECG electrodes could cause skin irritation; patients should be examined for signs of irritation or inflammation.
- To avoid the possibility of serious injury or death during patient defibrillation, do not come into contact with device or patient cables. Additionally, proper placement of defibrillator paddles in relation to the electrodes is required to minimize harm to the patient.
- Defibrillation protection is guaranteed only if the original patient cable is used. Any modification of this device may alter defibrillator protection.
- To avoid potential for spread of disease or infection, single-use disposable components (e.g., electrodes) must not be reused. To maintain safety and effectiveness, electrodes must not be used beyond their expiration date.
- This device was designed to use the electrodes specified in this manual. Proper clinical procedure must be employed to prep the electrode sites and to monitor the patient for excessive skin irritation, inflammation, or other adverse reactions. Electrodes are intended for short-term use and should be removed from the patient promptly following testing.

- FCC Warning (Part 15.21): Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the device.
- A possible explosion hazard exists. Do not use the device in the presence of a flammable anesthetic mixture.
- The device has not been designed for use with high-frequency (HF) surgical equipment and does not provide a protective means against hazards to the patient.
- The quality of the signal produced by the device may be adversely affected by the use of other medical equipment, including but not limited to defibrillators and ultrasound machines.
- There is no known safety hazard if other equipment, such as pacemakers or other stimulators, is used simultaneously with the device; however, disturbance to the signal may occur.
- Operations may be affected in the presence of strong electromagnetic sources such as electrosurgery equipment.
- The battery operated device transmits data reflecting a patient's physiological condition to a receiving device. During operation failure, data transmission and LCD information will cease to occur. In mission critical conditions, it is advisable to have a backup device available.
- Use only recommended alkaline battery cells. Use of other cells may present a risk of fire or explosion.
- Low battery warning function is designed for alkaline battery cells only. Use of other cells may result in failure of the low battery warning possibly resulting in a malfunction of the device.
- Do not attempt to clean the device or patient cables by submersing into a liquid, autoclaving, or steam cleaning as this may damage equipment or reduce its usable life. Use of unspecified cleaning/disinfecting agents, failure to follow recommended procedures, or contact with unspecified materials could result in increased risk of harm to users, patients and bystanders, or damage to the device. Do not sterilize the device or patient cables with Ethylene Oxide (EtO) gas.
- The ELI electrocardiograph does not automatically switch between the device and a direct patient cable. Clinician must choose the desired cable in the Settings menu before ECG acquisition. Switching between a direct patient cable and the device does not require the device to be re-paired with the ELI electrocardiograph unless a different device is to be used. The selected patient cable will display continuously under the Settings menu button.
- The minimum amplitude of the patient physiological signal that the ME equipment or system accepts is 30 μ V. Operation of the ME equipment or ME system below this amplitude may cause inaccurate results.



Caution:

- To prevent possible damage to the device, do not use sharp or hard objects to depress buttons, only use fingertips.
- The device and lead wires should be cleaned between each use. Inspect connections for damage or excessive wear prior to each use. Replace lead wires if damage or excessive wear is noted.
- Do not pull or stretch patient cables as this could result in mechanical and/or electrical failures. Lead wires should be stored after forming them into a loose loop.
- The device will only work with receiving devices that are equipped with the appropriate option.
- No user-serviceable parts are inside. Damaged or suspected inoperative equipment must be immediately removed from use and must be checked/repared by qualified service personnel prior to continued use.
- This device is not recommended for use in the presence of imaging equipment such as Magnetic Resonance Imaging (MRI) and Computed Tomography (CT) devices, etc.
- The following equipment may cause interference with the RF channel: microwave ovens, diathermy units with LANs (spread spectrum), amateur radios, and government radar.
- When necessary, dispose of the device, its components and accessories (e.g., batteries, cables, electrodes), and/or packing materials in accordance with local regulations.
- AA batteries are known to leak their contents when stored in unused equipment. Remove battery from device when not used for an extended period of time.
- Be careful to insert the connector block into the appropriate input connector by matching the lead wire labels to the device label.
- To prevent possible damage to the device during transport and storage (while in original packaging) the following environmental conditions must be adhered to:

Ambient Temperature Range: -20°C to 65°C (-4°F to 149°F)
Relative Humidity Range: 5% to 95% (non-condensing)
Atmosphere Pressure: 500 hPa to 1060 hPa

- This device is intended to be used in a hospital or doctor's office setting, and should be used according to the environmental conditions specified below:

Ambient Temperature Range: 0°C to 40°C (32°F to 104°F)
Relative Humidity Range: 5% to 95% (non-condensing)
Atmosphere Pressure: 500 hPa to 1060 hPa

Notes:

- Proper patient preparation is important to proper application of ECG electrodes and operation of the device.
- If electrode is not properly connected to the patient, or one or more of the patient cable lead wires is damaged, display will indicate a lead fault for the lead(s) where the condition is present.
- WAM™ (wireless acquisition module) must be paired to electrocardiograph before operation. If you possess more than one WAM, you are advised to label it and the paired electrocardiograph in order to avoid confusion. A sheet of labels is provided with the WAM for convenience.
- For additional instructions and warnings, refer to the user manual of the receiving device.
- As defined by IEC 60601-1 and IEC 60601-2-25, the device is classified as follows:
 - Class I equipment or internally powered.
 - Type CF (ECG) defibrillation-proof applied parts.
 - Equipment not suitable for use in the presence of a flammable anesthetic mixture.
 - Continuous operation.
- The device will automatically start flashing LEDs if the batteries have been discharged below 1.0 volts.
- During normal operation, the green LED will display continuously.
- If the battery cover is opened during transmission, the device will stop transmitting. The battery must be reinserted and the cover must be applied to resume operation.
- The device will automatically turn off (LEDs off) if the battery has been severely discharged.
- The device will automatically turn off when the electrocardiograph is powered down.
- The device will automatically turn off (LEDs off) after two minutes of inactivity or user input.
- Switching to a direct patient cable will automatically power down the device.
- When pairing with the device ensure the AM12 has been disconnected or the pairing operation will fail.
- To perform a STAT ECG using the device paired with an ELI 230 electrocardiograph, user must select STAT at the electrocardiograph to continue.
- A square wave presentation on the display while using the WAM may be due to the WAM being turned off, having no battery, not being paired correctly, operating out of range, or due to a calibration error. Review the LED indicator and auditory advisory on the WAM to ensure the unit is turned on, has proper battery level, is paired correctly, and is within recommended proximity of the electrocardiograph, or power cycle the WAM to re-calibrate.
- The device is UL classified:



WITH RESPECT TO ELECTRIC SHOCK,
FIRE AND MECHANICAL HAZARDS ONLY IN ACCORDANCE WITH
IEC 60601-1, CAN/CSA C22.2 No. 60601-1 and IEC 60601-2-25

EQUIPMENT SYMBOLS AND MARKINGS

Symbol Delineation

	<p>WARNING The warning statements in this manual identify conditions or practices that could lead to illness, injury, or death. In addition, when used on a patient applied part, this symbol indicates defibrillation protection is in the cables. Warning symbols will appear with a grey background in a black and white document.</p>
	<p>CAUTION The caution statements in this manual identify conditions or practices that could result in damage to the equipment or other property, or loss of data.</p>
	<p>Defibrillator-proof type CF applied part</p>
	<p>Acquire 12 Lead ECG</p>
	<p>Acquire Rhythm Print</p>
	<p>On / Off button</p>
	<p>Version 2 WAM indicator</p>
	<p>Version 2 UTK indicator</p>
	<p>Battery with notification: See operating instructions for type of cell to be used</p>
	<p>Separate collection of Electrical and Electronic Equipment. Do not dispose as unsorted municipal waste. Per European Union Directive 2012/19/EU, requires separate handling for waste disposal according to national requirements</p>
	<p>Indicates compliance to applicable European Union directives</p>
	<p>Non-ionizing electromagnetic radiation</p>
	<p>Follow instructions/directions for use (DFU) -- mandatory action. A copy of the DFU is available on this website. A printed copy of the DFU can be ordered from Hillrom for delivery within 7 calendar days.</p>



Medical Device



Product reference



Model Identifier




Lot Number



Serial Number



Manufacturer



Authorized representative in the European Community

Package Symbol Delineation



Keep away from sunlight



This way up



Fragile



Keep dry



Temperature limit



Humidity limitation



Atmospheric pressure limitation



Contains Non-spillable Battery

GENERAL CARE

Precautions

- Turn off the device before inspecting or cleaning.
- Do not immerse the device in water.
- Do not use organic solvents, ammonia-based solutions, or abrasive cleaning agents which may damage equipment surfaces.

Inspection

Inspect your equipment daily prior to operation. If you notice anything that requires repair, contact an authorized service person to make the repairs.

- Verify that all cables and connectors are securely seated.
- Check the case for any visible damage.
- Inspect cables and connectors for any visible damage.
- Inspect buttons and controls for proper function and appearance.

Cleaning and Disinfecting

Disinfecting agents

The WAM is compatible with the following disinfectants:

- Clorox Healthcare[®] Bleach Germicidal Wipes (use according to instructions on product label), or
- a soft, lint-free cloth dampened with a solution of sodium hypochlorite (10% household bleach and water solution) minimum 1:500 dilution (minimum 100 ppm free chlorine) and maximum 1:10 dilution as recommended by the APIC Guidelines for Selection and Use of Disinfectants.



Caution: Disinfecting or cleaning agents that contain Quaternary Ammonium Compounds (Ammonium Chlorides) have been identified as having negative effects if used to disinfect the product. Use of such agents may result in discoloration, cracking, and deterioration of the external housing of the device.

Cleaning

To clean the WAM:

1. Remove the battery.
2. Remove cables and lead wires from device before cleaning.
3. Thoroughly wipe the surface of the WAM with a clean, lint-free cloth dampened with a mild detergent and water for general cleaning or use one of the above recommended agents for disinfection.
4. Dry the device with a clean, soft, dry, lint-free cloth.

**WARNING:**

Prevent liquid from penetrating the device and do not attempt to clean/disinfect the device or patient cables by submerging into a liquid, autoclaving, or steam cleaning.

Do not expose cables to strong ultra-violet radiation.

Do not sterilize the device or lead wires with Ethylene Oxide (EtO) gas.

Do not immerse cable ends or lead wires; immersion can cause metal corrosion. Use caution with excess liquid as contact with metal parts may cause corrosion.

Do not use excessive drying techniques such as forced heat.

Improper cleaning products and processes can damage the device, produce brittle lead wires and cables, corrode the metal, and void the warranty. Use care and proper procedure whenever cleaning or maintaining the device.

Disposal

Disposal must be in accordance with the following steps:

1. Follow cleaning and disinfection instructions per instructions in this user manual section.
2. Segregate material in preparation for the recycling process
 - Components are to be disassembled and recycled based on type of material
 - Plastic to be recycled as plastic waste
 - Metal to be recycled as Metals
 - Includes loose components containing more than 90% metal by weight
 - Includes screws and fasteners
 - Electronic components, including the power cord, to be disassembled and recycled as Waste of Electrical and Electronic Equipment (WEEE)
 - Batteries to be dismantled from the device and properly disposed according with Battery directive.

Users must adhere to all federal, state, regional, and/or local laws and regulations as it pertains to the safe disposal of medical devices and accessories. If in doubt, the user of the device shall first contact Hillrom Technical Support for guidance on safe disposal protocols.



Waste of Electrical and
Electronic Equipment (WEEE)

INTRODUCTION

Manual Purpose

This manual is intended to provide the user with information about:

- Using and understanding the WAM (wireless acquisition module), the operator buttons, and the LED indicators.
- Preparing the WAM for use.
- Acquiring and printing an ECG.
- Maintenance.

NOTE: This manual may contain screen shots. Any screen shots are provided for reference only and are not intended to convey actual operating techniques. Consult the actual screen in the host language for specific wording.

Audience

This manual is written for clinical professionals who are expected to have a working knowledge of medical procedures and terminology as required for monitoring and/or acquiring clinical data from cardiac patients.

System Description

The WAM incorporates wireless electrocardiographic technology to achieve the acquisition and RF transmission of diagnostic-quality 12-lead ECG data. Transmission of the ECG data to a Welch Allyn receiver module allows the cardiac signals to be displayed on a monitoring device, such as an electrocardiograph, without the need for a direct connection.

The following equipment is necessary to use the WAM:

- One AA alkaline battery, 1.5V
- Electrocardiograph with Welch Allyn receiver module
- Lead wire sets
- ECG electrodes

Indications for Use

- Indicated for use as a radiofrequency physiological signal transmitter that acquires and delivers RF transmission of electrocardiographic data obtained during resting/physiologic electrocardiographic testing.
- The intended application, the patient population and the end use of the acquired data are determined by the host system. The Acquisition Modules are intended to be used by a licensed health care provider in a hospital or clinical setting.

WAM™ Wireless Acquisition Module and UTK Receiver




The USB Transceiver Key (UTK) is a compact radio transceiver designed to be interoperable with the WAM. The UTK is compatible with the USB ports of PCs and Welch Allyn ECG devices and enables these devices to receive, display, and store ECG data from patients with a wireless connection.

The UTK is internally built into the Welch Allyn ELI electrocardiographs.

WAM with Lead Wires

Figure 1-1



Using the Buttons

The WAM is operated by three buttons located on the front of the device:

- Power On/Off
- Acquiring a 12-lead ECG
- Acquiring a rhythm strip

Approved Battery Models

Description	Manufacturer	Part Numbers
Alkaline, AA-type, 1.5V	Various	Various



WARNING: Use of other cells may present a risk of fire or explosion.

To order additional supplies, contact a Welch Allyn customer service representative.

Part Numbers

Description	Part Numbers
Wireless Acquisition Module (WAM+) W/O LEAD WIRES	30012-019-56
UTK MODULE (receiver for WAM with Stress and Rscribe systems)	30012-021-51
UTK MODULE – 2 (receiver for WAM with Stress and Rscribe systems)	30012-021-54

WAM ACCESSORY KIT WITH AHA BANANA LEADS (includes WAM+)	41000-031-50
WAM ACCESSORY KIT WITH IEC BANANA LEADS (includes WAM+)	41000-031-51
WAM ACCESSORY KIT WITH AHA CLIP LEADS (includes WAM+)	41000-031-52
WAM ACCESSORY KIT WITH IEC CLIP LEADS (includes WAM+)	41000-031-53
BATTERY CAP ASSEMBLY WAM	8356-008-51
COMBINER WAM LEADS 10 POSITION IEC & AHA GRAY	9293-046-07
RPLCE LD SET WAM/AM12 FULL SET BANA AHA GRAY	9293-046-60
RPLCE LD SET WAM/AM12 FULL SET BANA IEC GRAY	9293-046-61
RPLCE LD SET WAM/AM12 LIMBS BANA AHA GRY	9293-046-62
RPLCE LD SET WAM/AM12 LIMBS BANA IEC GRY	9293-046-63
RPLCE LD SET WAM/AM12 V1-V3 BANA AHA GRY	9293-046-64
RPLCE LD SET WAM/AM12 C1-C3 BANA IEC GRY	9293-046-65
RPLCE LD SET WAM/AM12 V4-V6 BANA AHA GRY	9293-046-66
RPLCE LD SET WAM/AM12 C4-C6 BANA IEC GRY	9293-046-67
RPLCE LD SET WAM/AM12 FULL SET CLIP AHA GRAY	9293-047-60
RPLCE LD SET WAM/AM12 FULL SET CLIP IEC GRAY	9293-047-61
RPLCE LD SET WAM/AM12 LIMBS CLIP AHA GRY	9293-047-62
RPLCE LD SET WAM/AM12 LIMBS CLIP IEC GRY	9293-047-63
RPLCE LD SET WAM/AM12 V1-V3 CLIP AHA GRY	9293-047-64
RPLCE LD SET WAM/AM12 C1-C3 CLIP IEC GRY	9293-047-65
RPLCE LD SET WAM/AM12 V4-V6 CLIP AHA GRY	9293-047-66
RPLCE LD SET WAM/AM12 C4-C6 CLIP IEC GRY	9293-047-67
LEAD SET WAM/AM12 10-WIRE SHORT CLIPS AHA GRAY	9293-047-70
LEAD SET WAM/AM12 10-WIRE SHORT CLIPS IEC GRAY	9293-047-71
RPLCE LEAD SET WAM/AM12 LIMBS SHORT CLIPS AHA GRAY	9293-047-72
RPLCE LEAD SET WAM/AM12 LIMBS SHORT CLIPS IEC GRAY	9293-047-73
RPLCE LEAD SET WAM/AM12 V1-V3 SHORT CLIPS AHA GRAY	9293-047-74
RPLCE LEAD SET WAM/AM12 C1-C3 SHORT CLIPS IEC GRAY	9293-047-75
RPLCE LEAD SET WAM/AM12 V4-V6 SHORT CLIPS AHA GRAY	9293-047-76
RPLCE LEAD SET WAM/AM12 C4-C6 SHORT CLIPS IEC GRAY	9293-047-77
WAM USER MANUAL CD	9515-174-50-CD

WAM Specifications

Feature	Specification*
Instrument Type	12-lead wireless acquisition module for resting ECG
Input Channels	12-lead signal acquisition and transmission

ECG Leads Transmitted	I, II, III, aVR, aVL, aVF, V1, V2, V3, V4, V5, and V6
WAM Transmission Protocol	Bidirectional and frequency hopping; beacon and response method links a single acquisition module to a single electrocardiograph
Frequency Range	2403.38 MHz to 2479.45 MHz
Channel spacing	1MHz
RF output power	<10dBm
Antenna Type	PCB inverted F
Antenna Gain	-0.33dBi
Modulation	MSK
WAM and Receiver Distance	Approximately 10 feet (3 meters)
Lead Set	RA, LA, RL, LL, V1, V2, V3, V4, V5, and V6 (R, L, N, F, C1, C2, C3, C4, C5, and C6) with detachable lead wires
Sampling Rate	40,000 samples/second/channel acquisition; 1,000 samples/second/channel transmitted for analysis
Resolution	1.875 microvolt LSB
User Interface	Three-button operation: ON/OFF, 12-lead ECG acquisition, and rhythm strip acquisition
Defibrillator Protection	Complies with AAMI standards and IEC 60601-2-25
Special Functions	LED indication of power status, operating mode, lead fail, and remaining battery charge
Device Classification	Type CF, battery operated
Weight	6.7 oz. (190 g) with battery
Dimensions	4.45 x 4.25 x 1.1" (11.3 x 10.8 x 2.79 cm)
Battery	1 AA alkaline battery typically powers WAM for acquisition of 250 resting ECGs

**Specifications subject to change without notice*

UTK Specifications

Feature	Specification
Frequency	2403.38 ~ 2479.45MHz
Channel spacing	1MHz
RF output power	<10dBm
Antenna Type	PCB inverted F
Antenna Gain	-4.12dBi
Modulation	MSK

**Specifications subject to change without notice.*

EQUIPMENT PREPARATION

Battery Installation

The WAM is powered with a single AA battery. When the battery contains sufficient voltage to operate and the patient is properly connected, an LED on the front of the WAM will appear solid green indicating proper pairing and communication with the electrocardiograph. A battery with low voltage or a lead fail will result in a flashing green or yellow LED.

To install a new battery:

1. Remove the battery cover by twisting the cover in a counterclockwise direction.
 - a. Removal of the battery cover will automatically turn the power off.
2. Insert one AA battery into the battery compartment aligning the battery's positive (+) and negative (-) indicators with the designators shown on the device's back label.
3. Replace the battery cover by twisting the cover in a clockwise direction.
 - a. The battery cover will seal the battery compartment and make contact with the battery providing power to the device.

Applying Power

Before you apply power to the WAM, make sure that patient lead wires do not touch metal connected to ground (this may happen if reusable electrodes with exposed metal are used). The WAM will auto-calibrate at power-on, and a large amount of noise caused by ground loops may disrupt calibration, in which case the electrocardiograph will not display the ECG.

1. Press the Power On/Off button.
2. On initial startup, the LEDs will briefly flash yellow and green and the device will beep. LED status will indicate the following:

Status	Description
Solid green	Appropriate battery power level, good electrode-to-skin impedance, and good bidirectional communication with the electrocardiograph.
Flashing green	Low battery
Solid yellow	Lead fail
Flashing yellow	Low battery and/or lead fail
LED off	Device not powered on, very low battery (no sound), or device out of range (WAM will beep intermittently).

3. Press the Power On/Off button to turn the device off. An audible tone will sound indicating power off and RF disconnect.

Attaching the Lead Wire Connector Block

The 12-lead ECG lead wires consist of one connector block with 10 lead wires (5 lead wires to each side). The lead wires are positioned on the WAM to follow the contour of the torso. Each lead wire terminates in a medi-clip or 4 mm banana connector.

1. Securely insert the connector block into the ECG input connector on the top of the WAM.



CAUTION: Be careful to insert the connector block into the appropriate input connector by matching the lead wire labels to the WAM label.

Labeling the WAM and Electrocardiograph

The WAM ships with self-adhesive letters allowing the user to label the WAM and its paired electrocardiograph. It is recommended these labels are applied to both the WAM and its paired electrocardiograph to assist in keeping the units together.

Pairing with an ELI 150c/ELI 250c

Power on the ELI 150c or ELI 250c and:

1. Select **F6 MORE**.
2. Select **F6 MORE**.
3. Select **F2 WAM Pairing**.
4. Place the WAM (powered off) on top of the electrocardiograph.
5. Select **START**, then turn WAM on.
6. Enter password.
 - a. A successfully paired message will display.
7. Follow instructions on the display.

NOTE: When pairing with the WAM ensure the AM12 has been disconnected or the pairing operation will fail.

NOTE: The WAM will automatically turn off when the electrocardiograph is powered down.

Pairing with an ELI 230

Power on the ELI 230 and:

1. Select **MORE**.
2. Select **CONFIG**.
3. Select **4 Service**.
4. Select **Yes**.
5. Select **2 WAM Pairing**.
6. Place the WAM (powered off) on top of the ELI 230.
7. Select **START**, then turn WAM on.
 - a. A successfully paired message will display.
8. Select **DONE**.
9. Restart the ELI 230.

Pairing with an ELI 280

Power on the ELI 280 and:

1. Select **SETTINGS**.
2. Select **WAM**.
3. Place the WAM (powered off) on top of the ELI 280.
4. Select **Pairing**.
5. Select **START**, then turn WAM on.
 - a. A successfully paired message will display.
6. Select **DONE**.

Pairing with an ELI 350

The WAM will not take a 15-lead ECG. If the ELI 350 ECG Acquisition is set at 15 lead, the electrocardiograph will automatically switch to direct patient cable. To take a 12-lead ECG with the WAM the user must select 12 lead in the electrocardiograph's ECG Acquisition menu.

Power on the ELI 350 and:

1. Select **Settings** from the main display.
2. Enter password and select **WAM**.
 - a. Patient cable selection defaults to WAM. Remaining fields will automatically fill when pairing is done.
3. Select **Set Device ID**.
4. Select **Yes** to continue with pairing.
5. Per display instructions, turn off the WAM, place the WAM on top of the ELI 350, and turn the WAM on.
 - a. A successfully paired message will display.




WARNING: The ELI 350 does not automatically switch between the WAM and a direct patient cable. Clinician must choose the desired cable in the Settings menu before ECG acquisition. Switching between a direct patient cable and the WAM does not require the WAM to be re-paired with the ELI 350 unless a different WAM is to be used. The selected patient cable will display continuously under the Settings menu button.

- When the WAM is paired with an ELI 350 but the WAM is not on or its batteries are depleted, the ELI 350 will flash a “Searching for WAM” message under the Settings menu button.
- When the WAM is on and within range, the ELI 350 will display “WAM” and up to five signal bars. Also displayed is a battery gauge: green signifies adequate battery strength; red means the battery needs to be replaced immediately.
- When the WAM is on and within range but is not connected to a patient, “Leads Off” will display on the ELI 350 underneath heart rate. ECG leads will display as square waves.

NOTE: When using the ELI 350, switching to a direct patient cable will automatically power down the WAM.

NOTE: The WAM will automatically turn off when the electrocardiograph is powered down.

Pairing with an ELI 380

1. At the ELI 380 select  followed by **WAM/AM-XX**. Dependent on last saved setting, either AM12, AM15, or WAM is displayed with FPGA and UTK Firmware versions.
2. Select **Switch to WAM** followed by **WAM Pairing**. Follow the on-screen instructions. Once the WAM is paired, a Successfully Paired message will display.
3. Select **Done** to return to the Configuration screen.

Pairing the WAM with Q-Stress

Start the Q-Stress application. Start a stress test and navigate to the observation phase, then:

1. Select **Local Settings** and choose **WAM** as the Front End.
2. Select the **WAM Pairing** button.
3. Select **OK**.
4. Place the WAM (powered off) in close proximity to the UTK receiver connected to the Q-Stress USB port.
5. Turn WAM on.
6. A successfully paired message will display.
7. Select **OK**.

***NOTE:** Ending the stress exam will automatically cause the WAM to power off. It is not necessary to pair the WAM with the same UTK to use it again.*

***NOTE:** LED indication is not available when using WAM with Q-Stress.*

***NOTE:** 12-Lead ECG and Rhythm Print buttons are non-functional when using WAM with Q-Stress.*

Pairing the WAM with XScribe

Start the XScribe application. Start a stress test and navigate to the observation phase, then:

1. Select **Local Settings** and choose **WAM** as the Front End.
2. Select the **WAM Pairing** button.
3. Select **OK**.
4. Place the WAM (powered off) in close proximity to the UTK receiver connected to the XScribe USB port.
5. Turn WAM on.
6. A successfully paired message will display.
7. Select **OK**.


***NOTE:** Ending the stress exam will automatically cause the WAM to power off. It is not necessary to pair the WAM with the same UTK to use it again.*

***NOTE:** LED indication is not available when using WAM with XScribe.*

***NOTE:** 12-Lead ECG and Rhythm Print buttons are non-functional when using WAM with XScribe.*

WAM UTK Compatibility

A WAM with a “2” on its label will only be able to pair with a UTK that has a “2” on its label. Similarly, a WAM or UTK without a “2” will not be able to pair to a UTK or WAM that has a “2”. If there is trouble pairing the WAM, check the labels to ensure the WAM and UTK either both have a “2” or neither do.



Lead Fail

Lead fail is done automatically through visual communication with the LEDs located on the front of the WAM. A yellow LED (solid or flashing) indicates a lead fail condition is present. A solid green LED indicates proper lead connection as well as adequate battery voltage for ECG acquisition.

LED Indicators

LED	+ Audio	MODE
GREEN off YELLOW off	Intermittent beeping	Device is on but not paired to an electrocardiograph, is out of range of the paired electrocardiograph, or when used with ELI 350 could signify that direct patient cable is selected.
YELLOW solid or flashing GREEN off		One or more leads are not connected properly.
GREEN solid YELLOW off		No lead fail condition is detected; battery is OK.
GREEN solid YELLOW off	Intermittent beeping	Device is collecting a 10-second ECG.
Blinking LED (yellow or green depending on lead fault status)		Device has detected a low battery condition. Replace the battery within 15 minutes.
GREEN off YELLOW off	1 second audio on, then device turns off.	Device has detected a very low battery status and powered off.

ACQUIRING AN ECG

Use the LED indicators to check electrode-to skin impedance and verify patient hookup quality, as well as to ensure communication has been established with the electrocardiograph and the signal quality of each ECG is transmitted as expected. A yellow LED indicates a lead fail condition.

1. Ensure an AA battery is in the battery compartment. If battery voltage is too low, the WAM may not power on. Insert a new AA battery into the device to continue operation.
2. Press the On/Off button to turn the WAM on.
3. Connect the patient to the WAM lead wires (*see Patient Hookup in the recording device user manual*).
4. ECG data should be automatically transmitted to the recording device.
5. Enter patient information at the recording device.
6. Press the 12-Lead ECG Acquisition button to complete the acquisition of the 12-lead ECG.
7. Press the Rhythm Print button to acquire a rhythm print; press the Rhythm Print button again to stop the rhythm print.

NOTE: During normal operation, the green LED will display continuously.

NOTE: If the battery cover is opened during transmission, the WAM will stop transmitting. The battery must be reinserted and the cover must be applied to resume operation.

NOTE: Using the buttons to acquire a 12-lead ECG or a rhythm strip is not functional with XScribe.

8. At the end of the ECG acquisition session, the WAM should be turned off. ECG data may now be reviewed, plotted, or edited as needed at the electrocardiograph

NOTE: To perform a STAT ECG using the WAM paired with an ELI 230 electrocardiograph, user must select STAT at the electrocardiograph to continue.

MAINTENANCE

Periodic Maintenance

Check the WAM and lead wires before each use to ensure they are not damaged or broken.

MESSAGES AND INFORMATION

The following table defines LED signals seen at the WAM during patient hookup and transmission.

Message	Solution
LED off, intermittent beeping	WAM is on but not synched to an electrocardiograph. Ensure that the electrocardiograph is turned on. Follow the pairing process if necessary.
Solid green LED	No interaction required.
Flashing green LED	Replace AA battery.
Solid yellow LED	Lead fail message, check leads for proper connection.
Flashing yellow LED	Low battery and lead fail condition exist. Replace battery and check leads for proper connection.
LED Off, no beeping	Power is off, press power button to turn on. If no audible beeping is heard, battery is completely depleted. Replace battery to power on the WAM. If beeping is heard ensure you are within 10 ft. (3 meters) of the paired electrocardiograph.

The following messages are seen on the ELI electrocardiograph where applicable. NOTE: lead wire messages will be seen on the ELI electrocardiograph display and not on the WAM; the WAM will display a yellow LED in the event of a lead failure.

Message	Solution
RA or N	RA or N fail. Check if the lead wire is off or the electrode needs to be replaced.
RL or R	RL or R fail. Check if the lead wire is off or the electrode needs to be replaced.
LA or F	LA or F fail. Check if the lead wire is off or the electrode needs to be replaced.
LL or L	LL or L fail. Check if the lead wire is off or the electrode needs to be replaced.
A combination of RA/.../LL or N...F	More than one limb lead fail or all leads fail. Check the lead wires and electrodes.
V1 or C1	V/C1 fail. Check if the lead wire is off or the electrode needs to be replaced.
V2 or C2	V/C2 fail. Check if the lead wire is off or the electrode needs to be replaced.
V3 or C3	V/C3 fail. Check if the lead wire is off or the electrode needs to be replaced.
V4 or C4	V/C4 fail. Check if the lead wire is off or the electrode needs to be replaced.
V5 or C5	V/C5 fail. Check if the lead wire is off or the electrode needs to be replaced.
V6 or C6	V/C6 fail. Check if the lead wire is off or the electrode needs to be replaced.
A combination of V1, V2, V3, V4, V5, V6, or C1, C2, C3, C4, C5, C6	More than one chest lead fails. Check the lead wires and electrodes.

The following system information log is provided for your convenience. You need this information if your device needs servicing. Be sure to update the information log when you add options or your device has been serviced.

Record the model and serial number of all components, dates of removal and/or replacement, and the name of the vendor from whom the component was purchased and/or installed.

In addition to having records of this information, the system information log provides a warranty record of when your device was placed in service.

System Information Log

Manufacturer:

Welch Allyn, Inc.
4341 State Street Road
Skaneateles Falls, NY 13153

Telephone Numbers:

USA: 800-231-7437
Sales Department: 800-231-7437
Service Department: 888-667-8272

Product Information:**Name of Unit/Product:** _____**Date of Purchase:** ___/___/___**Purchased Unit From:**
_____**Serial Number:** _____**Software Version:** _____

Serial and Part Number Location

For questions and service information, have both serial and part number available when calling.

The model type, serial number (SN), and part number (REF) are found on the back label of the device as described in the Notices section of this manual.

ELECTROMAGNETIC COMPATIBILITY (EMC)

Electromagnetic compatibility with surrounding devices should be assessed when using the device.

An electronic device can either generate or receive electromagnetic interference. Testing for electromagnetic compatibility (EMC) has been performed on the device according to the international standard for EMC for medical devices (IEC 60601-1-2). This IEC standard has been adopted in Europe as the European Norm (EN 60601-1-2).

The device should not be used adjacent to, or stacked on top of other equipment. If the device must be used adjacent to or stacked on top of other equipment, verify that the device operates in an acceptable manner in the configuration in which it will be used.

Fixed, portable, and mobile radio frequency communications equipment can affect the performance of medical equipment. See appropriate EMC table for recommended separation distances between the radio equipment and the device.

The use of accessories, transducers, and cables other than those specified by Welch Allyn may result in increased emissions or decreased immunity of the equipment.

Guidance and Manufacturer's Declaration: Electromagnetic Emissions

The equipment is intended for use in the electromagnetic environment specified in the table below. The customer or the user of the equipment should ensure that it is used in such an environment.

Emissions Test	Compliance	Electromagnetic Environment: Guidance
RF Emissions CISPR 11	Group 1	The equipment must emit electromagnetic energy in order to perform its intended function. Nearby electronic equipment may be affected. The equipment is suitable for use in all establishments including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
RF Emissions CISPR 11	Class B	
Harmonic Emissions IEC 61000-3-2	Not Applicable	
Voltage Fluctuations/ Flicker Emissions IEC 61000-3-3	Not Applicable	

Guidance and Manufacturer's Declaration: Electromagnetic Immunity


The equipment is intended for use in the electromagnetic environment specified in the table below. The customer or the user of the equipment should ensure that it is used in such an environment.

Emissions Test	Compliance	Compliance Level	Electromagnetic Environment: Guidance
Electrostatic discharge (ESD) IEC 61000-4-2	+/- 6 kV contact +/- 8 kV air	+/- 6 kV contact +/- 8 kV air	Floors should be wood, concrete, or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst IEC 61000-4-4	+/- 2 kV for power supply lines +/- 1 kV for input/output lines	Not Applicable	
Surge IEC 61000-4-5	+/- 1 kV differential mode +/- 2 kV common mode	Not Applicable	
Voltage dips, short interruptions, and voltage variations on power supply input lines IEC 61000-4-11	<5% UT (>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles	Not Applicable	
Power frequency (50/60 Hz) magnetic field	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.

NOTE: UT is the AC Mains voltage prior to application of the test level.

Guidance and Manufacturer's Declaration: Electromagnetic Immunity

The equipment is intended for use in the electromagnetic environment specified in the table below. The customer or the user of the equipment should ensure that it is used in such an environment.

Emissions Test	IEC 60601 Test Level	Compliance Level	Electromagnetic Environment: Guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz	3 Vrms 150 kHz to 80 MHz	<p>Portable and mobile RF communications equipment should be used no closer to any part of the equipment, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended separation distance</p> $d = \left[\frac{3.5}{3V_{rms}} \right] \sqrt{P}$ $d = \left[\frac{3.5}{3V/m} \right] \sqrt{P} \quad 80 \text{ MHz to } 800 \text{ MHz}$ $d = \left[\frac{7}{3V/m} \right] \sqrt{P} \quad 800 \text{ MHz to } 2.5 \text{ GHz}$ <p>Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m).</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey^a, should be less than the compliance level in each frequency range^b.</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p> 
Radiated RF IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m 80 MHz to 2.5 GHz	

NOTE 1: At 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by the absorption and reflection from structures, objects, and people.

- a. Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radios, AM and FM radio broadcast, and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the equipment is used exceeds the applicable RF compliance level above, the equipment should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the equipment.
- b. Over the frequency range 150 kHz to 80 MHz, field strengths should be less than [3] V/m.

Recommended Separation Distances Between Portable and Mobile RF Communications Equipment and the Equipment

The equipment is intended for use in the electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the equipment can help to prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the equipment as recommended in the table below, according to the maximum output power of the communications equipment.

Rated Maximum Output Power of Transmitter W	Separation Distance According to Frequency of Transmitter (m)		
	150 kHz to 80 MHz	80 MHz to 800 MHz	800 MHz to 2.5 GHz
	$d = 1.2\sqrt{P}$	$d = 1.2\sqrt{P}$	$d = 2.3\sqrt{P}$
0.01	0.1 m	0.1 m	0.2 m
0.1	0.4 m	0.4 m	0.7 m
1	1.2 m	1.2 m	2.3 m
10	4.0 m	4.0 m	7.0 m
100	12.0 m	12.0 m	23.0 m

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1: At 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by the absorption and reflection from structures, objects, and people.

Regulatory Radio Compliance

Federal Communications Commission (FCC)

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- This device may not cause harmful interference.
- This device must accept any interference received, including interference that may cause undesired operation.

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy. If not installed and used in accordance with the instructions, it may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try and correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna
- Increase the distance between the equipment and the receiver
- Connect the equipment to an outlet on a circuit different from that to which the receiver is connected
- Consult the dealer or an experienced radio/TV technician for help

The user may find the following booklet prepared by the Federal Communications Commission helpful: The Interference Handbook This booklet is available from the U.S. Government Printing Office, Washington, D.C. 20402. Stock No. 004-000-0034504. Welch Allyn is not responsible for any radio or television interference caused by unauthorized modification of the devices included with this Welch Allyn product, or the substitution or attachment of connecting cables and equipment other than specified by Welch Allyn. The correction of interference caused by such unauthorized modification, substitution, or attachment will be the responsibility of the user.

WAM : HJR-WAM2500 UTK : HJR-UTK2500
--

Industry Canada (IC) Emissions

RF Radiation Hazard Warning

Using higher gain antennas and types of antennas not certified for use with this product is not allowed. The device shall not be co-located with another transmitter.

Cet avertissement de sécurité est conforme aux limites d'exposition définies per la norm CNR-102 at relative aux fréquences radio.

This device complies with RSS 210 of Industry Canada.

Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of this device.

L'utilisation de ce dispositif est autorisée seulement aux conditions suivantes: (1) il ne doit pas produire de brouillage et (2) l'utilisateur du dispositif doit être prêt à accepter tout brouillage radioélectrique reçu, même si ce brouillage est susceptible de compromettre le fonctionnement du dispositif.

This Class B digital apparatus complies with Canadian ICES-003.

Cet appareil numérique de la classe B est conform à la norme NMB-003 du Canada.

WAM : 3758B-WAM2500 UTK : 3758B-UTK2500
--

Declaración de conformidad Mexico





La operación de este equipo está sujeta a las siguientes dos condiciones:

1. es posible que este equipo o dispositivo no cause interferencia perjudicial y
2. este equipo o dispositivo debe aceptar cualquier interferencia, incluyendo la que pueda causar su operación no deseada.

European Union

Czech	Welch Allyn tímto prohlašuje, že tento WLAN device je ve shodě se základními požadavky a dalšími příslušnými ustanoveními směrnice 2014/53/ES.
Danish	Undertegnede Welch Allyn erklærer herved, at følgende udstyr WLAN device overholder de væsentlige krav og øvrige relevante krav i direktiv 2014/53/EF
Dutch	Bij deze verklaart Welch Allyn dat deze WLAN device voldoet aan de essentiële eisen en aan de overige relevante bepalingen van Richtlijn 2014/53/EC.
English	Hereby, Welch Allyn, declares that this WLAN device is in compliance with the essential requirements and other relevant provisions of Directive 2014/53/EC.
Estonian	Käesolevaga kinnitab Welch Allyn seadme WLAN device vastavust direktiivi 2014/53/EÜ põhinõuetele ja nimetatud direktiivist tulenevatele teistele asjakohastele sätetele.
Finnish	Welch Allyn vakuuttaa täten että WLAN device tyyppinen laite on direktiivin 2014/53/EY oleellisten vaatimusten ja sitä koskevien direktiivin muiden ehtojen mukainen.
French	Par la présente, Welch Allyn déclare que ce WLAN device est conforme aux exigences essentielles et aux autres dispositions de la directive 2014/53/CE qui lui sont applicables
German	Hiermit erklärt Welch Allyn die Übereinstimmung des Gerätes WLAN device mit den grundlegenden Anforderungen und den anderen relevanten Festlegungen der Richtlinie 2014/53/EG. (Wien)
Greek	ΜΕ ΤΗΝ ΠΑΡΟΥΣΑ Welch Allyn ΔΗΛΩΝΕΙ ΟΤΙ WLAN device ΣΥΜΜΟΡΦΩΝΕΤΑΙ ΠΡΟΣ ΤΙΣ ΟΥΣΙΩΔΕΙΣ ΑΠΑΙΤΗΣΕΙΣ ΚΑΙ ΤΙΣ ΛΟΙΠΕΣ ΣΧΕΤΙΚΕΣ ΔΙΑΤΑΞΕΙΣ ΤΗΣ ΟΔΗΓΙΑΣ 2014/53/ΕΚ
Hungarian	Alulírott, Welch Allyn nyilatkozom, hogy a WLAN device megfelel a vonatkozó alapvető követelményeknek és az 2014/53/EC irányelv egyéb előírásainak.
Italian	Con la presente Welch Allyn dichiara che questo WLAN device è conforme ai requisiti essenziali ed alle altre disposizioni pertinenti stabilite dalla direttiva 2014/53/CE.
Latvian	Ar šo Welch Allyn deklarē, ka WLAN device atbilst Direktīvas 2014/53/EK būtiskajām prasībām un citiem ar to saistītajiem noteikumiem.
Lithuanian	Šiuo Welch Allyn deklaruoja, kad šis WLAN device atitinka esminius reikalavimus ir kitas 2014/53/EB Direktyvos nuostatas.
Malti	Hawnhekk, Welch Allyn, jiddikjara li dan WLAN device jikkonforma mal-htigijiet essenzjali u ma provvedimenti oħrajn relevanti li hemm fid-Direttiva 2014/53/EC
Portuguese	Welch Allyn declara que este WLAN device está conforme com os requisitos essenciais e outras disposições da Directiva 2014/53/CE.
Slovak	Welch Allyn týmto vyhlasuje, že WLAN device spĺňa základné požiadavky a všetky príslušné ustanovenia Smernice 2014/53/ES.
Slovene	Šiuo Welch Allyn deklaruoja, kad šis WLAN device atitinka esminius reikalavimus ir kitas 2014/53/EB Direktyvos nuostatas.
Spanish	Por medio de la presente Welch Allyn declara que el WLAN device cumple con los requisitos esenciales y cualesquiera otras disposiciones aplicables o exigibles de la Directiva 2014/53/CE
Swedish	Härmed intygar Welch Allyn att denna WLAN device står i överensstämmelse med de väsentliga egenskapskrav och övriga relevanta bestämmelser som framgår av direktiv 2014/53/EG.

Radio Compliance Table for:

Argentina	Ente Nacional de las Comunicaciones (ENACOM)	 COMISIÓN NACIONAL DE COMUNICACIONES H-22661 (WAM) H22622 (UTK)	
Australia	Australian Communications and Media Authority (ACMA) Radio Compliance Mark (RCM).		
Brazil	Agência Nacional de Telecomunicações (ANATEL)	 Modelo: WAM 01142-14-05187	Este equipamento opera em caráter secundário, isto é, não tem direito a proteção contra interferência prejudicial, mesmo de estações do mesmo tipo, e não pode causar interferência a sistemas operando em caráter primário.
EAC			Products meet all requirements of the corresponding technical regulations and have passed all conformity assessment procedures.
Indonesia	<p>Keterangan</p> <p>a. [58975/SDPPI/2018] (UTK) adalah nomor sertifikat yang diterbitkan untuk setiap alat dan perangkat telekomunikasi</p> <p>b. [8260] (UTK) adalah nomor PLG ID (identitas pelanggan) berdasarkan database Lembaga Sertifikasi</p> <p>a. [60825/SDPPI/2019] (WAM) adalah nomor sertifikat yang diterbitkan untuk setiap alat dan perangkat telekomunikasi</p> <p>b. [8260] (WAM) adalah nomor PLG ID (identitas pelanggan) berdasarkan database Lembaga Sertifikasi</p>	<p>Identification</p> <p>a. [58975/SDPPI/2018] (UTK) is a number of certificate issued for certified telecommunication equipment</p> <p>b. [8260](UTK) is a number of PLG ID based on one Certification Body database</p> <p>a. [60825/SDPPI/2019] (WAM) is a number of certificate issued for certified telecommunication equipment</p> <p>b. [8260](WAM) is a number of PLG ID based on one Certification Body database</p>	
Mexico	Instituto Federal de Telecomunicaciones (Federal Telecommunications Institute—IFETEL)	<p>This product contains and Approved module, Model No. WAM, IFETEL No. RCPWEWA19-0527</p> <p>This product contains and Approved module, Model No. UTK, IFETEL No. RCPWEUT19-0526</p>	La operación de este equipo está sujeta a las siguientes dos condiciones: (1) es posible que este equipo o dispositivo no cause interferencia perjudicial y (2) este equipo o dispositivo debe aceptar cualquier interferencia, incluyendo la que pueda causar su operación no deseada.
Morocco		AUTHORIZED BY MOROCCO ANRT WAM: Approval number: MR 17489 ANRT 2018	

Date of approval: 13-SEP-2018

UTK: Approval number: MR 17488 ANRT 2018

Date of approval: 13-SEP-2018

Oman	Telecommunications Regulatory Authority		WAM : R/6168/18 D172250 UTK: R/6164/18 D172250
Paraguay	Comisión Nacional de Telecomunicaciones		NR:121/2019 (WAM) NR:122/2019 (UTK)
Pakistan	Pakistan Telecom Authority		
Philippines	National Telecommunications Commission		WAM: ESD-18-18399C UTK: ESD-19-19449C
Singapore	Info-Communications Media Development Authority (IMDA)		
South Korea	Korea Communications Commission (대한민 국 방송통 신위원 회) – KCC Certification number: WAM: IC MSIP-CRI-S83-WAM UTK: IC MSIP-CRI-S83-UTK		This equipment is Industrial (Class A) electromagnetic wave suitability equipment and seller or user should take notice of it, and this equipment is to be used in the places except for home. 이 기 기 는 업 무 용 (A 급) 전 자 파 적 합 기 기 로 서 판 매 자 또는 사 용 자 는 이 점 을 주 의 하 시 기 바 라 며, 가 정 외 의 지 역 에 서 사 용 하 는 것 을 목 적 으 로 합 니 다.
UAE			WAM: ER65767/18 UTK: ER65804/18

Class A Equipment (Industrial Broadcasting & Communication Equipment)

A급 기기(업무용 방 송통신기자재)